

HIGH ACCURACY MEASUREMENTS OF RINGS, PLUGS, STEP-GAUGES AND BALL-BARS USING A CMM COUPLED WITH AN INTERFEROMETRIC LASER

JOÃO ANTÔNIO PIRES ALVES, MSC

*Instituto Nacional de Metrologia, Normalização e Qualidade Industrial - Inmetro
Duque de Caxias, RJ, 20250-020, Brazil
jaalves@inmetro.gov.br*

JOSELAINÉ ROSSIGNOLI ZANETTI

*Instituto Nacional de Metrologia, Normalização e Qualidade Industrial - Inmetro
Duque de Caxias, RJ, 20250-020, Brazil
jrzanetti@inmetro.gov.br*

Campus of Inmetro

Materials

Mechanic

Operacional

Thermal


Electrical

Chemistry


Acoustic

And Vibration

Optics



STRUCTURE



LAMED-DIMENSIONAL METROLOGY LABORATORY

INMETRO/LAMED - SERVICES

- **Gage – blocks, bars;**
- **Roughness and depth set standards;**
- **Angular blocks, polygons;**
- **Roundness standards;**
- **Laser Systems for linear measurements;**
- **Step-gages, ball-bars, 2D artefacts;**
- **Rings, plugs and spheres;**
- **Scales;**



**Industries
and
accredited
labs**

TOTAL : 26 SERVICES

CMM FACILITIES AT LAMED



Room 1

LEGEX -9106 – MITUTOYO

- # 2D ARTEFACTS**
- # STEP-GAGES BY
COMPARISON**
- #GENERAL MEASUREMENTS**



Room 2

UMM 500 – ZEISS

- # RINGS, PLUGS, SPHERES**
- # STEP-GAGES, BALL-BARS**
- #GENERAL MEASUREMENTS**

STEP-GAGES CALIBRATION




Gage-blocks

Step-gage

Gages positioning

Set-up

Position of the
sensors on the
gage-blocks



Position of the
sensors on the step-
gage

20mm, 50mm,
100m, 150mm,...
600mm

Calibration strategy

1 – Probing error determination



**2 – Measurements of blocks.
Smaller >>> Larger**




3 – Step-gage measurement





5 cycles

Measurement uncertainty




BALL - PLATE CALIBRATION



Gage-blocks


HOLE - PLATE CALIBRATION



Gage-blocks

Measurement uncertainty - Ball-plate


x_i	$u(x_i)$	ν_i	$c_i = \partial l / \partial x_i$	$u_i(l) / \text{nm}$
Plate measurements				
Short Term Reproducibility (Plate)	0,173205081	Infinite	1000	173,2
Thermometer calibration	0,002	Infinite	11,5	0,02
Thermal Expantion Coeficient (Plate)	1,1547E-06	Infinite	125330	0,1
Gradients and Instability (Plate)	0,025980762	Infinite	11,5	0,3
Comparison				
Short Term Reproducibility (Plate)	0,083715789	infinite	1000	83,7
Reproducibility	0,111407653	4	1000	111,4
Gage-blocks calibration	0,03	Infinite	1000	30,0
Repeatability - Blocks	0,193304565	4	1000	193,3
Repeatability - Plate	0,117004273	4	1000	117,0
Thermal Expantion Coeficient (Plate and Gage-blocks)	1,1547E-06	Infinite	59170	0,1
Gradients and Instability (gage-blocks)	0,067261306	Infinite	11,5	0,8
Gradients and Instability (Plate)	0,068704682	Infinite	11,5	0,8



Combined standard uncertainty: $u_c(l) = \sqrt{(318,397)^2 + (1,15671)^2 \cdot l^2}$ nm; l , in mm.

Measurement uncertainty - Hole-plate

x_i	$u(x_i)$	ν_i	$c_i = \frac{\partial l}{\partial x_i}$	$u_i(l) / \text{nm}$
Plate measurements	---	---	---	---
Short Term Reproducibility (Plate)	0,53982	Infinite	1000	539,82
Thermometer calibration	0,00200	Infinite	0,50	0,00
Thermal Expansion Coefficient (Plate)	0,00000	Infinite	236166,67	0,07
Gradientes and Instability (Plate)	0,11172	Infinite	0,50	0,05
Comparison	---	---	---	---
Short Term Reproducibility (Blocks)	0,173205	infinite	1000	173,20
Reproducibility	0,339664	4	1000	339,66
Gage-blocks calibration	0,065	Infinite	1000	65
Repeatability - Blocks	0,230478	4	1000	230,47
Repeatability - Plate	0,259808	4	1000	259,81
Thermal Expansion Coefficient (Plate and Gage-blocks)	1,15E-06	Infinite	145333,33	0,17
Gradientes and Instability (gage-blocks)	0,110274	Infinite	11,5	1,27
Gradientes and Instability (Plate)	0,121532	Infinite	11,5	1,40



Combined standard uncertainty: $u_c(l) = \sqrt{(749,416)^2 + (1,89671)^2} \cdot l$ nm; l , in mm.

Comparison with pilot results- Ball plate

Esfera No.	INMETRO		NMIJ		INMETRO		Emx	Emy
	DX [mm]	DY [mm]	U_x [um]	U_y [um]	U_x [um]	U_y [um]		
1	0,0000	0,0000	0,12	0,12	0,75	0,75	0,00	0,00
2	0,0002	-0,0001	0,15	0,12	0,76	0,75	0,26	-0,08
3	0,0001	0,0000	0,20	0,12	0,77	0,75	0,10	-0,03
4	0,0002	-0,0001	0,27	0,12	0,80	0,75	0,29	-0,07
5	0,0002	0,0000	0,34	0,12	0,84	0,75	0,17	0,00
6	-0,0001	0,0001	0,12	0,15	0,89	0,75	-0,09	0,14
7	0,0002	0,0002	0,15	0,15	0,94	0,75	0,18	0,32
8	0,0001	0,0001	0,20	0,15	1,00	0,75	0,09	0,13
9	0,0001	0,0001	0,27	0,15	1,07	0,75	0,10	0,14
10	0,0001	0,0001	0,34	0,15	1,14	0,75	0,07	0,15
11	0,0000	0,0002	0,12	0,20	1,21	0,75	-0,04	0,20
12	0,0001	0,0002	0,15	0,20	1,28	0,75	0,10	0,22
13	0,0001	0,0001	0,20	0,20	0,75	0,76	0,08	0,09
14	0,0002	0,0001	0,27	0,20	1,28	0,76	0,12	0,13
15	0,0001	0,0001	0,34	0,20	0,75	0,77	0,12	0,18
16	-0,0001	0,0000	0,12	0,27	1,28	0,77	-0,07	-0,03
17	0,0002	0,0001	0,15	0,27	0,75	0,80	0,21	0,09
18	0,0001	-0,0001	0,20	0,27	1,28	0,80	0,06	-0,16
19	0,0002	0,0000	0,27	0,27	0,75	0,84	0,26	0,03
20	0,0002	0,0000	0,34	0,27	1,28	0,84	0,14	-0,01
21	-0,0002	0,0001	0,12	0,34	0,75	0,89	-0,21	0,11
22	0,0001	0,0002	0,15	0,34	1,28	0,89	0,07	0,19
23	0,0000	0,0002	0,20	0,34	0,75	0,94	0,00	0,17
24	0,0002	0,0001	0,27	0,34	1,28	0,94	0,12	0,14
25	0,0001	0,0001	0,34	0,34	0,75	1,00	0,14	0,09

$$En_{x,y} = D_{x,y} / \sqrt{(U^2 + Ur^2)}$$

Maximum value


0,32

CMM – LASER SYSTEM




UMM 500 – ZEISS

Environmental conditions- Room 2 – long period




11h00min

Environment condition - Room 2 – short period




12h00min

LENGTH MEASUREMENT SYSTEM



Error sources



Refraction index of the air

$$\lambda_{\text{air}} < \lambda_{\text{vac}}$$

$$\lambda_{\text{vac}} = \eta \lambda_{\text{ar}}$$

η Refraction index of the air
 λ_{vac} wave-length in vacuun;
 λ_{air} wave-length in air;

Refraction index of the air - Edlén :


$$\eta = \left[p \times \frac{(\eta_s - 1)}{K_1} \right] \times \frac{\left(+10^{-8} \times (K_2 + K_3 T) \times p \right)}{(1 + K_4 T)} - f \times \left[K_5 + K_6 \times \left(\frac{1}{\lambda} \right)^2 \right] \times 10^{-10} + 1$$

where :

- p Air pressure - Pa;
- f Partial pressure for moist air - Pa;
- T Air temperature - °C;

K_i are constants

Dead-path




$$DP = L1 * \alpha * (\Delta T)$$

“Cossine error”

L = Laser displacement


M = Table displacement



$$\xi = \frac{M - L}{M} = 1 - \cos \theta = \frac{\theta^2}{2}$$

“Cossine error”

$$\xi = \frac{S^2}{8L^2}$$



Gage temperature

$$L_{20} = L_e / (1 + \alpha * (T_e - 20))$$


α = Thermal expansion coefficient;

L_e = Length at environmental temperature;

L_{20} = Length at 20 °C;

T_e = Calibration temperature;

Ideal model




Abbè principle




Only if in laser direction

Set up - Alignment and fixation of the gage




Calibration based in KOBA geometry

Horizontal positioning



Alignment error



Error $\Delta Z \sim 5\mu\text{m}$


Error $DZ \sim 1\mu\text{m}$

- Actual model


$$l_g = l_l \pm d_p - E_{la} - E_{ga} - E_{dl} - E_{sys}$$

$$l_g = \bar{l}_l * \frac{n_0}{n} * \frac{1}{(1 + \alpha * (T_g - T_r))} \pm \frac{d_p}{(1 + \alpha * (T_g - T_r))} - E_{la} - E_{ga} - E_{dl} - E_{sys}$$

Gage-block measurement



Systematic errors



Using the calibrated system

Step gage calibration

$$0\text{mm} - 420\text{mm} - (0,4^2 + (1,2 * 10^{-3} * (L))^2)^{0,5}$$

$$420\text{mm} - 880\text{mm} - 0,7 + (0,4^2 + (1,2 * 10^{-3} * (L - 420))^2)^{0,5}$$

$$880\text{mm} - 1060\text{mm} - 1,4 + (0,4^2 + (1,2 * 10^{-3} * (L - 880))^2)^{0,5}$$

L in mm

Using the calibrated system

One single client


Tree-spheres Ball-bar

Sphere - 1 to 2 = 221mm


Sphere – 2 to 3 = 311mm

$$0mm - 420mm - (0,2^2 + (1,2 * 10^{-3} * (L))^2)^{0,5}$$


Step gages – Nominal Length up to 420mm




420mm < Nominal Length < 620mm



$620\text{mm} < \text{Nominal Length} < 1060\text{mm}$



Errors in overlapping




DATA-SHEET – CONSTANT PART

3 - Planilha de incerteza								
Grandezas de Influência	Estimativa	Distribuição	Divisor	Incert. Padrão [μm]	Ci	Contrib. ui [μm]	ui ² [μm] ²	vi
Contribuições independentes do comprimento								
Resolução	[μm] 0,0100	Retangular	3,4641	0,003	1	0,0029	0,0000083	
Instabilidade	[μm] 0,0100	Retangular	1,73205	0,006	1	0,00577	0,0000333	
Alinhamento Laser	[μm] 0,0313	Retangular	1,73205	0,018	1	0,01804	0,0003255	
Padrão	0,0100	Retangular	1,73205	0,006	1	0,00577	0,0000333	
Comprimento morto	[μm] 0,0230	Retangular	1,73205	0,013	1	0,01328	0,0001763	
Apalpador	[μm] 0,2	Normal	2	0,075	1	0,07500	0,0056250	
						Total 1 =	0,0062019	


DATA-SHEET – LENGTH DEPENDENT PART

3 - Planilha de incerteza								
Grandezas de Influência	Estimativa	Distribuição	Divisor	Incert. Padrão [μm]	Ci	Contrib. ui [μm]	ui ² [μm] ²	vi
Temperatura do Ar	[°C]							
Incerteza Certificado	0,0200000	Normal	2	0,010	-9,36E-07	-9,357E-09	8,756E-17	
Oscilação no tempo	0,0280000	Retangular	1,73205	0,016	-9,36E-07	-1,513E-08	2,288E-16	
Pressão Atmosférica	[Pa]							
Incerteza Certificado	52,5956743	Normal	2	26,298	2,69E-09	7,065E-08	4,991E-15	
Oscilação no tempo	8,2627871	Retangular	1,73205	4,771	2,69E-09	1,282E-08	1,642E-16	
Umidade do ar	[Pa]							
Incerteza Certificado	22,9063111	Normal	2	11,453	-3,70616E-10	-4,245E-09	1,802E-17	
Oscilação no tempo	8,2462720	Retangular	1,73205	4,761	-3,70616E-10	-1,764E-09	3,113E-18	
Comprimento de onda	[μm]							
	0,00000007	Normal	2	0,000	-2,19557E-06	-7,684E-14	5,905E-27	
						Total 2 =	5,493E-15	
Temperatura do Padrão	[°C]							
Incerteza Certificado	0,0200000	Normal	2	0,010	1,15E-05	1,150E-07	1,323E-14	
Oscilação no tempo	0,0040000	Retangular	1,73205	0,002	1,15E-05	2,656E-08	7,053E-16	
Coef. de dilatação	[1 /°C]							
	0,000001	Retangular	1,73205	0,000	-0,349	-2,015E-07	4,060E-14	
						Total 3 =	5,453E-14	

Comparison with BMC




BALL - BARS




Beginning of the calibrated range

Probing the spheres




BALL - BARS



Beginning of the calibrated range

Probing the spheres



$$BMC = (0,2^2 + (1,2 * 10^{-3} * L)^2)^{1/2}, L \text{ in mm}$$

Rings, plugs and spheres calibration

- >Results of calibrated system not used**
- >Evaluation of the probing errors each calibration**
- >Differences between internal and external probing**
- >Measurements in the centre of the machine table (Range : 200mm)**

BMC – Same as ball-bars

Trouble

Trouble : Only 4 gages


>11mm and 52mm rings

>One single 30mm sphere


>One single 50mm plug

Solution: Gage blocks


The right measurement strategy




Gage-block - Strategy 1



Ring - gage - Strategy 2



Ring-gage - Strategy 3



Ring-gage - Strategy 4

Strategies 1 and 3 lead to same results

Some text results

Gage-blocks

Probing error (mm)	Date	Result for gage length (mm)	Difference from the certificate (mm)
4,99747	03/09/2009	49,99997	0,02
4,99747	04/09/2009	49,99995	0,00

Two spheres

Probing error (mm)	Date	Result for gage length (mm)	Difference from the certificated (mm)
4,99737	26/08/2009	19,98553	0,00
4,99727	26/08/2009	19,98560	0,07
4,99731	27/08/2009	19,98557	0,04
4,99731	27/08/2009	19,98555	0,02
4,99741	28/08/2009	19,98540	-0,13
4,99735	28/08/2009	19,98561	0,03

Gage-block and ring-gage

Probing error (mm)	Date	Result for gage length (mm)	Difference from the certificated (mm)
4,99747	03/09/2009	51,99930	-0,10
4,99752	04/09/2009	51,99940	0,00
4,99756	04/09/2009	51,99953	0,13
4,99754	04/09/2009	51,99944	0,04
4,99751	04/09/2009	51,99939	-0,01

Gage-block and plug

Probing error (mm)	Date	Result for gage length (mm)	Difference from the certificate (mm)
4,99764	26/08/2009	50,00073	-0,07
4,99769	27/08/2009	50,00045	-0,35
4,99765	27/08/2009	50,00059	-0,21
4,99763	28/08/2009	50,00058	-0,22

Some clients results


Ball-bar

Ball-bar - RC 1794				
Certificate	221mm	$U_{221} (\mu m)$	311mm	$U_{311} (\mu m)$
1343/2009	221,0047	0,5	311,0050	0,6
1959/2004	221,0048	0,5	311,0050	0,6

Ring-gage

Ring gage			
Serial Nbr	Identification	Model	
1165	IMT - 023	Ident. Nbr.	
Certificate			
0301/2003		0544/2005	
V.V.C (mm)	U (μm)	V.V.C (mm)	U (μm)
49,9996	0,2	49,99954	0,21
49,9995	0,2	49,99955	0,21
Certificate			
0963/2007		1286/2009	
V.V.C (mm)	U (μm)	V.V.C (mm)	U (μm)
49,9994	0,3	49,9994	0,3
49,9995	0,3	49,9995	0,3

Step-gage



Conclusions

System demonstrated good performance

*Comparisons: EUROMET.L-K5.2004 – Step Gauge (420mm), EUROMET.L-K4.2005
– Diameter Standards, APMP.L-K5.2006 - Step Gauge (620mm).*

But

Now under a re-evaluation.

Calibration with more points - 0mm to 200mm.

Ball-bar calibrations – client asks for better uncertainties.

NACMA – 3

3rd Tri-National Conference of the North American Coordinate Metrology Association



Ministério do
Desenvolvimento, Indústria
e Comércio Exterior



Thank you!
Muchas gracias!

Questions and suggestions are welcome!
